

fatigue vs PBO after 6 months of treatment. Reduced severity in insomnia symptoms with LEM5 and LEM10 also translated to improved daytime functioning. Funding Acknowledgements: Supported by Eisai Inc.

166

Post Market Rate of Seizures During TMS Treatment with NeuroStar® System Appears to Be Lower than Previously Estimated

Paul Janicak, MD¹; Karen Heart, MBA²; and Bridget McGugan, PharmD, MBA³

¹ Psychiatrist, Rush University Medical Center, Rush University, Naperville, IL

² Director, Medical Operations, Neuronetics, Inc., Malvern, PA

³ Clinical Research Manager, Medical Operations, Neuronetics, Inc., Malvern, PA

ABSTRACT: Objective: NeuroStar® Advanced Therapy System is a transcranial magnetic stimulation (TMS) device with FDA-clearance for the treatment of Major Depressive Disorder (MDD) in adult patients who have failed to receive satisfactory improvement from prior antidepressant medication in the current episode. With TMS, magnetic pulses are transmitted into the brain. Though the exact mechanism of action is unknown, it is postulated that resulting neuronal depolarization and changes in brain functional activity may be associated with various physiologic changes that lead to relief of depression in the indicated population. The type of magnetic field generated with TMS is not intended to induce a seizure during therapeutic use, but unintentional seizures have been reported during TMS treatment.

No seizures were reported with the use of the NeuroStar® system in clinical trials conducted prior to FDA clearance. The estimated risk of seizure in the NeuroStar® label is approximately 1 in 30,000 treatments or 1 in 1,000 patients. Since introduction of the NeuroStar® system into clinical practice, the rate at which seizures have been reported is even lower.

METHODS: We conducted a review of literature that named the NeuroStar® Advanced Therapy System as the device used for TMS treatment and reviewed all seizure events reported to Neuronetics, Inc., directly or through FDA MedWatch through June 30, 2019. Articles reporting seizures in subjects with epilepsy during TMS treatment were excluded.

RESULTS: Previous comprehensive reviews of seizures induced by treatment with any TMS device by Wasserman et al. (1998) and Rossi et al. (2009) revealed that the rate of seizures is low. Many subjects that developed seizures during TMS had either received stimulation at

parameters beyond current recommendations or had been predisposed to develop seizures in some way. Some of the events reported as seizures may, in fact, have been non-epileptic events.

Our literature review and analysis of seizures reported to Neuronetics, Inc. revealed that the rate of seizures during TMS treatment with the NeuroStar® appears to be lower than the rate that is published in the NeuroStar® Advanced Therapy prescribing information.

CONCLUSIONS: Seizures that take place during TMS treatment with the NeuroStar® system are rare. The rate of seizures reported directly to Neuronetics, Inc. is lower than that included in the NeuroStar® prescribing information. Our literature review validated seizures during TMS treatment with the NeuroStar® system reported in published literature have described either non-epileptic events (syncope) or occurred with risk factors for seizure induction, such as other predisposing clinical factors or treatment parameters outside the guideline recommend “safe” ranges.

Funding Acknowledgements: Neuronetics, Inc.

167

Randomized, Double-Blind, Active-Controlled Study of Starting Aripiprazole Lauroxil with 1-Day Initiation in Acutely Ill Patients with Schizophrenia

Peter J. Weiden, MD¹; Amy Claxton, PhD²; Yangchun Du, PhD³; Sergey Yagoda, MD, PhD⁴; David Walling, PhD⁵; and Jelena Kunovac, MD⁶

¹ Senior Medical Director, Medical Affairs, Alkermes, Inc., Waltham, MA

² Director, Clinical Research, Alkermes, Inc., Waltham, MA

³ Senior Director, Biostatistics, Alkermes, Inc., Waltham, MA

⁴ Associate Medical Director, Clinical Research, Alkermes, Inc., Waltham, MA

⁵ CEO and Principal Investigator, CNS Network, LLC, Garden Grove, CA

⁶ CEO and CMO, Altea Research, Las Vegas, NV

ABSTRACT: Objective: Evaluate efficacy and safety of a 2-month dose of aripiprazole lauroxil (AL) with a 1-day initiation regimen during hospitalization for an acute exacerbation of schizophrenia.

METHODS: In the phase 3b double-blind ALPINE study, adults with schizophrenia were randomized to AL (AL NanoCrystal® Dispersion + oral aripiprazole 30 mg day 1; AL 1064 mg day 8 and every 8 weeks) or paliperidone palmitate (PP 234 mg day 1; PP 156 mg day 8 and every 4 weeks). Patients were discharged after 2 weeks of hospitalization and followed through week 25. Primary